



Ref. Certif. No.

DE 2-018119

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE  
CERTIFICAT D'ESSAI OC

Product  
Produit

AC/DC Switching Adaptor

Name and address of the applicant  
Nom et adresse du demandeur

Mean Well Enterprises Co.,Ltd.  
No. 28, Wuquan 3rd Rd.,  
248 Wugu Dist., New Taipei City, Taiwan

Name and address of the manufacturer  
Nom et adresse du fabricant

Mean Well Enterprises Co.,Ltd.  
No. 28, Wuquan 3rd Rd.,  
248 Wugu Dist., New Taipei City, Taiwan

Name and address of the factory  
Nom et adresse de l'usine

Mean Well Enterprises Co.,Ltd.  
No. 28, Wuquan 3rd Rd.,  
248 Wugu Dist., New Taipei City, Taiwan

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

AC 100-240V; 50/60Hz; 1.4-0.7A; Class II;  
for output ratings refer to test report no. 11037279 001  
page 5

Trade mark (if any)  
Marque de fabrique (si elle existe)

Trademark of MEAN WELL

Model/type Ref.  
Ref. de type

GSM120Bx  
(x= 12, 15, 20, 24 or 48 to denote different output voltage)

Additional information (if necessary may also be reported on page 2)  
Les Information complémentaire (si nécessaire, peuvent être indiqués sur la 2<sup>ème</sup> page)

Models differ in model name, output rating, windings of main transformer T1 and the ratings of components; the risk management requirements of the standard were not addressed; for details see test report ref. no. 11037279 001

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1:2005 + A1  
IEC 60601-1-11:2010  
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue une partie de ce Certificat

11037279 002

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com



Date: 15.09.2014

Signature:

Dipl.-Ing. W. Hsu