

Verification of Compliance

Product Name : AC/DC SWITCHING ADAPTOR
Brand Name : MEAN WELL
Model Number : GSM40Ax(x=05, 07, 09, 12, 15, 18, 24, 48),
GSM60Ax(x=05, 07, 09, 12, 15, 18, 24, 48), GSM40A48-CIS
Applicant : MEAN WELL ENTERPRISES CO., LTD.
Address : No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,
Taiwan (R.O.C.)
Report Number : S4C-M020-1302-120
Issue Date : December 19, 2016
Applicable Standards :

Emission :

EN 55011 : 2009+A1 : 2010

Group 1, Class B

EN 61000-3-2 : 2014

EN 61000-3-3 : 2013

EN 61204-3:2000 Class B

Immunity :

EN 60601-1-2 : 2015 Home Healthcare Environment

EN 61204-3 : 2000 : Low severity level

EN 61000-4-2 : 2009

EN 61000-4-3 : 2006+A1 : 2008+A2 : 2010

EN 61000-4-4 : 2012

EN 61000-4-5 : 2014

EN 61000-4-6 : 2014

EN 61000-4-8 : 2010

EN 61000-4-11 : 2004

Based on the EMC Directive 2014/30/EU, Medical Device Directive 93/42/EEC and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



TAF 0905

FCC CAB Code TW1053, TW0905, TW1104

NVLAP Lab Code 200575-0

IC Code 4699A

VCCI Accep. No. R-1527, C-1609, T-1441, G-10,
C-4400, T-1334, G-614



Central Research Technology Co.

EMC Test Laboratory

11, Lane41, Fushuen St., Jungshan Chiu,
Taipei, Taiwan, 104, R.O.C.

Tel : 886-2-25984568

Fax: 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: December 19, 2016