

Certificate



Certificate no.

TA 50218063 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10034229 002

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: MPx-yz (MW)	8
x stands for	: S, D, T	1
y stands for	: 45 or 65	1
z stands for	: -3.3, -5, -7.5, -12, -13.5, -15, -24, -27, -48, A, B or C	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1) 1.6A for MPx-65z 2) 1.2A for MPx-45z	
Output ratings	: see constructional dataform	
max. Ambient Temperature	: 60°C	
Protection Class	: I	


Remark: This license sheet replaces certificate TA 50218063, sheet 01 dated 30.11.2011; The risk management requirements of the standard were not addressed.

This certificate is valid for maximum 5 years.

11

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

08/12/2011

Certificate



Certificate no.

TA 50218063 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-SCL 10034229 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01


Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

30/11/2011

Certificate



Certificate no.

TA 50218063 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-SCL 10034229 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01


Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

30/11/2011

Certificate



Certificate no.

TA 50218063 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10034229 003

Client Reference: 05082014_A11

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01


Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

22/08/2014

Certificate



Certificate no.

TA 50218063 05

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10034229 004

Client Reference: FL-20150706-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

26/10/2015



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50218063 06

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI- 10034229 005

Client Reference: ES-18042017-TUV/MPx-

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above
Product : see above
License Holder : see above
Factory : see above

Remark: 1. The risk management requirements of the standard
were not addressed;
2. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

10/04/2018

Certificate



Certificate no.

TA 50218063 07

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road, Dong-qiao, Pan-
yang Ind. Park, Huang-dai Town, Xiang-
cheng District Suzhou, Jiangsu 215152
P. R. China

Test report no.: ZTW1- 10034229 005

Client Reference: ES-18042017-TUV/MPx-

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Change

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

10/04/2018

Certificate



Certificate no.

TA 50218063 08

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town, Huadu District,
Guangzhou 510890
P. R. China

Test report no.: ZTW1- 10034229 005

Client Reference: ES-18042017-TUV/MPx-

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Type Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above
max. Operational Altitude : 3000m

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

10/04/2018