

Certificate



Certificate no.

TA 50293226 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: RPX-60Y (MW)	10
X stands for	: S, D, T	1
Y stands for	: -3.3, -5, -12, -15, -24, -24CA, -48, A, B, C, D, 03	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.8-1.0A for X = S 1.4-0.9A for X = D, T	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature	: 40°C	
Protection Class	: I	

Continued on page 02

12

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue
(day/mo/yr)

16/09/2014

Certificate



Certificate no.

TA 50293226 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Continuation

Remarks:

1. The risk management requirements of the standard were not addressed.
2. Model RPS-60-24CA is evaluated for operating at altitude 6000 m. Other models are evaluated for operating at altitude 3000 m.
3. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
16/09/2014

Certificate



Certificate no.

TA 50293226 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: RPS-40-18.2 (MW)	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.4-0.8A	
Output Voltage	: DC +18.2V	
Output Current	: 2.2A	

This certificate is valid for maximum 5 years.

1

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
16/09/2014

Certificate



Certificate no.

TA 50293226 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
16/09/2014

Certificate



Certificate no.

TA 50293226 05

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

16/09/2014

Certificate



Certificate no.

TA 50293226 06

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Youngyear Electronics Co., Ltd.
Jin Zhu Industries Area
Qing Xi Town, Dongguan City
Guandong
P.R. China

Test report no.: ZTWI-SCL 10048037 001

Client Reference: 1407035002-003/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)
16/09/2014

Certificate



Certificate no.

TA 50293226 07

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10048037 002

Client Reference: FL-20150706-A12

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

27/10/2015



Dipl.-Ing. (FH) A. Klinker