



Ref. Certif. No.

DE 2-021998

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC

Product
Produit

AC/DC Medical Adaptor

Name and address of the applicant
Nom et adresse du demandeur

MEAN WELL Enterprises Co.,Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248, Taiwan

Name and address of the manufacturer
Nom et adresse du fabricant

MEAN WELL Enterprises Co.,Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248, Taiwan

Name and address of the factory
Nom et adresse de l'usine

☒ Additional Information page 2

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

AC 100-240V; 50/60Hz; 2.0-1.0A; Class I;
for output ratings refer to report no. 50062869 001
page 8

Trade mark (if any)
Marque de fabrique (si elle existe)

Trademark of MEAN WELL

Model/type Ref.
Ref. de type

GSM160Ax
(x = 12, 15, 20, 24 or 48 to denote the output voltage,
current and power ratings)

Additional information (if necessary may also be
reported on page 2)
Les Information complémentaire (si nécessaire,
peuvent être indiqués sur la 2^{ème} page)

Models differ in model name, output ratings, windings of
main transformer T1 (different numbers of secondary windings,
gauge, turns for different output ratings) and the ratings
of components; for the details see test report

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

PUBLICATION IEC 60601-1:2005 + A1
IEC 60601-1-6:2010 + A1
IEC 62366:2007 + A1
for national deviations see test report

EDITION


As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue une partie de ce Certificat

50062869 003

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 30.06.2017

Signature:

Dipl.-Ing. W. Hsu



Ref. Certif. No.

DE 2-021998

1. MEAN WELL Enterprises Co.,Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

2. Suzhou MEAN WELL Technology
Co., Ltd. No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng Dist.
Suzhou, 215152 Jiangsu, China

3. MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road
Huadong Town, Huadu District
Guangzhou, Guangdong, China

Additional information (if necessary)
Information complémentaire (si nécessaire)

Date: 30.06.2017

Signature:

Dipl.-Ing. W. Hsu

