

Verification of Compliance

Product Name : Power Supply
Brand Name : MEAN WELL
Model Number : MSP-300-x (x=3.3,5,7.5,12,15,24,36,48)
Applicant : MEAN WELL ENTERPRISES CO., LTD.
Address : No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,
Taiwan (R.O.C.)
Report Number : S2C-M020-1110-171
Issue Date : December 19, 2016
Applicable Standards :

Emission :

EN 55011 : 2009+A1 : 2010 Group 1, Class B
EN 61000-3-2 : 2014
EN 61000-3-3 : 2013

Immunity :

EN 60601-1-2 : 2015 Home Healthcare Environment
EN 61000-4-2 : 2009
EN 61000-4-3 : 2006+A1 : 2008+A2 : 2010
EN 61000-4-4 : 2012
EN 61000-4-5 : 2014
EN 61000-4-6 : 2014
EN 61000-4-8 : 2010
EN 61000-4-11 : 2004

Based on the EMC Directive 2014/30/EU, Medical Device Directive 93/42/EEC and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



TAF 0905

FCC CAB Code TW1053, TW0905, TW1104

NVLAP Lab Code 200575-0

IC Code 4699A

VCCI Accep. No. R-1527, C-1609, T-1441, G-10,
C-4400, T-1334, G-614



Central Research Technology Co.

EMC Test Laboratory
11, Lane41, Fushuen St., Jungshan Chiu,
Taipei, Taiwan, 104, R.O.C.
Tel : 886-2-25984568
Fax: 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: December 19, 2016