



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

Technical Standard: MDD DIRECTIVE 93/42/EEC

General Information

Applicant: N/A

Product Description

EUT Description: AC/DC SWITCHING ADAPTER

Brand Name: MEAN WELL

Model Number: GSM90Ax (x=12,15,19,24,48)

Measurement Standard

EN 55011:2009/A1:2010(CISPR 11:2009/A1:2010)

EN 61000-3-2: 2014

EN 61000-3-3: 2013

EN 60601-1-2:2015

EN 61000-4-2:2009; EN 61000-4-3:2010

EN 61000-4-4:2012 ; EN 61000-4-5:2014 ; EN 61000-4-6:2014

EN 61000-4-8:2010 ; EN 61000-4-11:2004

Measurement Facilities

Kunshan BU. Name: **Compliance Certification Services Inc.**

10#Weiye Rd, Innovation Park, Eco. & Tec, Development Zone, Kunshan city,
JiangSu, (215300) China.

Tel: + 86-512-57355888/ FAX: +86-512-57370818

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number:

C160202E02-ET

Jeff Fang

Jeff.Fang /EMC Manager

Date: March 03, 2016



TESTING CERT #2541.01

CCSRF 程智电子科技(昆山)有限公司
Compliance Certification Services(KunShan)Inc.