

Certificate



Certificate no.

TA 50308353 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-LinPr 10050387 001

Client Reference: 1T-Mark_PM-10, NFM-1

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: PM-10-X (MW)	8
X stands for	: 3.3, 5, 12, 15, 24	1
Rated Voltage	: 100-240V, 50/60Hz	
Rated Current	: 0.25-0.15A	
Outputs	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature	: 50°C	
Protection Class	: II	

Remarks: 1. The risk management requirements of the standard
were not addressed;
2. This certificate is valid for maximum 5 years.

9

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/04/2015

Certificate



Certificate no.

TA 50308353 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-LinPr 10050387 001

Client Reference: 1T-Mark_PM-10, NFM-1

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: NFM-10-X (MW)	1
X stands for	: 3.3, 5, 12, 15, 24	1

This certificate is valid for maximum 5 years.

Appendix: 1

2


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

10/04/2015

Certificate



Certificate no.

TA 50308353 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-LinPr 10050387 001

Client Reference: 1T-Mark_PM-10, NFM-1

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/04/2015

Certificate



Certificate no.

TA 50308353 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-LinPr 10050387 001

Client Reference: 1T-Mark_PM-10, NFM-1

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

10/04/2015