

VERIFICATION OF COMPLIANCE

Issue Date: August 18, 2021
Applicant: MEAN WELL ENTERPRISES CO., LTD.
Address: No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan (R.O.C.)
Manufacturer / Address: 1. Mean Well Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan (R.O.C.)
2. MEAN WELL(GUANGZHOU) ELECTRONICS CO., LTD.
No.11 Jingu South Road, Huadong Town, Huadu District, Guangzhou, China.
3. SuZhou Mean Well Technology Co., Ltd.
No.77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District, Suzhou, Jiangsu 215152, P.R. China
4. YONGDEN TECHNOLOGY CORPORATION
345 MacArthur Highway Tabang, Guiguinto, Bulacan, Philippines
5. MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED
9C, Peenya industrial area, Chokkasandra, 2ND PHASE, PEENYA, Bengaluru (Bangalore) Urban, Karnataka, 560058
Contact Information:
Product: Medical Type Switching Power Supply
Brand Name/Trade Mark: MEAN WELL
Model/Type: RPS-200-12, RPS-200-15, RPS-200-24, RPS-200-27, RPS-200-48
Added Model(s): RPS-200-X(X=12,15,24,27,48)
Applicable Standards: EN 55011: 2016/A11: 2020
EN IEC 61000-3-2: 2019+A1:2021
EN 61000-3-3: 2013/A1: 2019
EN 60601-1-2: 2015
EN 61000-4-2: 2009
EN 61000-4-3: 2006/A2: 2010
EN 61000-4-4: 2012
EN 61000-4-5: 2014/A1: 2017
EN 61000-4-6: 2014
EN 61000-4-8: 2010
EN IEC 61000-4-11: 2020
Test Laboratory: Compliance Certification Services Inc.
Wugu Lab
No.11, Wugong 6th Rd., Wugu Dist.,
New Taipei City, Taiwan. (R.O.C.)
Test Report No. : T210707W01-E, dated on August 18, 2021
Conclusion: Based upon a review of the Test Report(s), the tested sample of the product mentioned above is deemed to comply with the requirements of the above standards.
Note: This verification is only valid for the product and configuration described and in conjunction with the test report as detailed above.

Authorised Signatory:



Compliance Certification Services Inc.
Hex Chiang
Supervisor