

Certificate



Certificate no.

TA 50268288 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

ZTW1-LinPr-10038200 001

Test report no.: ZTW1-LinPr 10038200 001

Client Reference: HH110913

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Medical Power Adaptor)

License Fee - Units

Type Designation	: GSM06Ey-zzz (MW)	5
y stands for	: 05, 06, 07, 09, 12, 15, 18, 24	1
-zzz stands for	: 0-9, A-Z or blank	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.18-0.09A	
Output Voltage	: see Appendix	
Output Current	: see Appendix	
max. Power Output	: see Appendix	
max. Ambient Temperature	: 40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.

7

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)
23/10/2013

Certificate



Certificate no.

TA 50268288 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

ZTW1-LinPr-10038200 001

Test report no.: ZTW1-SCL 10038200 002

Client Reference: 22.07.2014

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Medical Power Adaptor)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu

Date of Issue
(day/mo/yr)

01/08/2014



Appendix to TÜV approved Certificate No.: TA 50268288

Certified Product : Medical Power Adaptor

Report Number : 10038200 001

Type Designation: GSM06Ey-zzz

(y=05, 06, 07, 09, 12, 15, 18, 24; -zzz=0-9, A-Z or blank)

Mean Well	Output Voltage [V]	Output Current [A]	Max. Power [W]
GSM06Ey-zzz			
GSM06E05-P1J	5	0-1.2	6
GSM06E06-P1J	6	0-1	6
GSM06E07-P1J	7.5	0-0.8	6
GSM06E09-P1J	9	0-0.66	6
GSM06E12-P1J	12	0-0.5	6
GSM06E15-P1J	15	0-0.4	6
GSM06E18-P1J	18	0-0.33	6
GSM06E24-P1J	24	0-0.25	6

Date: 23rd Oct., 2013

Certification Body


Dipl.-Ing. W. Hsu

Certificate



Certificate no.

TA 50268288 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

ZTW1-LinPr-10038200 001

Test report no.: ZTW1- 10038200 003

Client Reference: GW-27072015-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Medical Power Adaptor)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

26/11/2015