



Ref. Certif. No.

DK-5350

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Switching Power Supply

Name and address of the applicant
Nom et adresse du demandeur

MEAN WELL ENTERPRISES CO LTD
28 WUQUAN 3RD RD WUGU DIST
NEW TAIPEI, 248 Taiwan

Name and address of the manufacturer
Nom et adresse du fabricant

MEAN WELL ENTERPRISES CO LTD
28 WUQUAN 3RD RD WUGU DIST
NEW TAIPEI, 248 Taiwan

Name and address of the factory
Nom et adresse de l'usine

See Page 2

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

100-240 V ac, 50/60 Hz, 1.8-1.1A
Output: See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur
Model / Type Ref.
Ref. De type

RPx-75y (x can be S, D, T for number of output; y can be -3.3, -5, -12, -15, -24, -36 or -48, when x = S; y is to denote different output voltage; when x= D; y can be A or B, when x = T; y can be A, B, C, D or 03)

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{eme} page

Built-in

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);Corr.:1, IEC 60601-1(ed.3);Corr.:2

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

1104036003 issued on 2011-12-06

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com
Signature:

Date: 2011-12-09

Jan-Erik Storgaard

Factories:

Mean Well Enterprises Co., Ltd.
 No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei
 City 248, Taiwan

Mean Well (GUANGZHOU) Electronics Co., Ltd.
 2nd Floor, No. A Building, Yuean Ind. Park,
 Dongpu Town, TianHe Dist, Guangzhou, China

SuZhou Mean Well Technology Co., Ltd.
 No. 77, Jian-min Road, Dong-qiao, Pan-yang
 Ind. Park, Huang-dai Town, Xiang-cheng District,
 Suzhou, Jiangsu 215152, China

Output ratings:

Model name	Output rating	Output power	Transformer
RPS-75-3.3	3.3V/15A	49.5	TF-1531
RPS-75-5	5V/14A	70.0	TF-1532
RPS-75-12	12V/6.3A	75.6	TF-1533
RPS-75-15	15V/5A	75.0	TF-1534
RPS-75-24	24V/3.2A	76.8	TF-1535
RPS-75-36	36V/2.1A	75.6	TF-1536
RPS-75-48	48V/1.6A	76.8	TF-1537
RPD-75A	5V/7.0A, 12V/3.0A	71.0	TF-1522
RPD-75B	5V/5A, 24V/2.0A	73.0	TF-1523
RPT-75A	5V/6.0A, 12V/3.0A, -5V/0.5A	68.5	TF-1524
RPT-75B	5V/6.0A, 12V/3.0A, -12V/0.5A	72.0	TF-1525
RPT-75C	5V/6.0A, 15V/2.3A, -15V/0.5A	72.0	TF-1526
RPT-75D	5V/5.0A, 24V/1.5A, 12V/1.0A	73.0	TF-1527
RPT-7503	3.3V/6.0A, 5V/6.0A, 12V/1.0A	61.8	TF-1528

Additional information (if necessary)
Information complémentaire (si nécessaire)



UL International Demko A/S, Borupvang 5A, DK-2750
 Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
 www.ul-europe.com

Date: 2011-12-09

Signature:



Jan-Erik Storgaard