

Certificate



Certificate no.

TA 50219942 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-BHU 10034234 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: PM-15-X (MW)	8
X stands for	: 3.3, 5, 12, 15, 24 for different output voltage	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 0.35A-0.2A	
Output Voltage	: DC 3.3V, 5.0V, 12.0V, 15.0V, 24.0V	
Output Current	: 3.5A, 3.0A, 1.25A, 1.0A, 0.63A	
max. Ambient Temperature	: 40°C	
Protection Class	: II	

Remark: The risk management requirements of the standard were not addressed.

This certificate is valid for maximum 5 years.

Appendix: 1

9


Licensed Test mark:

TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue
(day/mo/yr)

02/01/2012



Dipl.-Ing. A. Klinker



Certificate



Certificate no.

TA 50219942 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTWI-BHU 10034234 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: NFM-15-X (MW)	1
X stands for	: 3.3, 5, 12, 15, 24 for different output voltage	1

This certificate is valid for maximum 5 years.

Appendix: 1

2

Licensed Test mark:


TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

02/01/2012



Dipl.-Ing. A. Klinker



Certificate



Certificate no.

TA 50219942 03

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well (GUANGZHOU) Electronics
Co Ltd
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-BHU 10034234 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01


Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. A. Klinker

Date of Issue
(day/mo/yr)

02/01/2012



Certificate



Certificate no.

TA 50219942 04

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou Mean Well Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-BHU 10034234 001

Client Reference: 8208400364/JH110811

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:

TÜV Rheinland Taiwan Ltd.


Signature

Date of Issue

(day/mo/yr)

02/01/2012




Dipl.-Ing. A. Klinker



Certificate



Certificate no.

TA 50219942 05

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-SCL 10034234 002

Client Reference: 05082014_A11

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)
22/08/2014

Certificate



Certificate no.

TA 50219942 06

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10034234 003

Client Reference: FL-20150706-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

26/10/2015



Dipl.-Ing. (FH) A. Klinker