

Certificate



Certificate no.

TA 50348281 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-LinPr 10054494 001 **Client Reference:** FL_02102015_RPS-200I

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

Type Designation	: RPS-200-x (MW)	10
x stands for	: 12, 15, 24, 27 or 48	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 2.3-1.1A	
Outputs	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature	: 50°C	
Protection Class	: II	
max. Operational Altitude	: 5000m	

Remarks: 1) The risk management requirements of the standard were not addressed;
2) This certificate is valid for maximum 5 years.

11

Appendix: 1

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419036526

TÜV Rheinland Taiwan Ltd

Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

14/06/2016

Certificate



Certificate no.

TA 50348281 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuguan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1-LinPr 10054494 001 **Client Reference:** FL_02102015_RPS-200I

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

14/06/2016

Certificate



TÜV Rheinland

This page was cancelled due to factory changed.
Please check the last page for description.

Certificate no.

TA 50348281 03

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) Electronics
Co., Ltd.
2nd Floor, No. A Building,
Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou
P.R. China

Test report no.: ZTW1-LinPr 10054494 001 **Client Reference:** FL_02102015_RPS-200I

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419036526

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

14/06/2016

明緯企業股份有限公司
MEAN WELL Enterprises Co., Ltd.

Mr. Johnny Huang, Safety & EMC Engineer
R & D Dept.

248 新北市五股區五權三路 28 號
No. 28, Wuquan 3rd Rd.
Wugu Dist., New Taipei City 248
Taiwan

Date: 7 November,
2016
Client No.: 331921
Our ref.: RNK/EVC

Ref.: Change of Factory Data

Dear Mr. Huang,

Please find the acknowledgement of the change of factory data for the existing product certificates from TÜV Rheinland Taiwan Ltd.

Delete factory:

MEAN WELL (GUANGZHOU) Electronics Co., Ltd.
2nd Floor, No. A Building, Yuean Ind. Park, Dongpu Town,
TianHe Dist., Guangzhou, P. R. China

Additional factory:

**MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD
HUADU BRANCH**
No. 11 Jingu South Road, Huadong Town,
Huadu District, Guangzhou 510890, P.R. China

Please notes that the updated factory data will be indicated in the product certificate to be issued, started from date of above; but the existing product certificates are not to be revised with this regards.

Sincerely yours,

TÜV RHEINLAND TAIWAN LTD.
Power Supply & Display Product Line


Robert Kong
Department Manager

TÜV Rheinland Taiwan Ltd.

11F., No.758, Sec. 4, Bade
Rd., Songshan Dist., Taipei
City 105, Taiwan R.O.C.
Tel. (02) 2172 7000
Fax (02) 2528 0018
<http://www.twn.tuv.com>

TAICHUNG BRANCH:
No. 9, Ln 36, Minsheng Rd.,
Sec. 3, Daya District,
Taichung City 428,
Taiwan, R.O.C.
Tel. (04) 2560 2998
Fax (04) 2566-3598

Customer Hotline:
(02)2528 7007
www.tuv.com

MEAN WELL Enterprises Co., Ltd.
Mr. Frank Lin Safety Engineer

Date : 08.02.2021
Our ref. : ZTW1
Your ref.: P00158594

No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Ref : TA License

Type of Equipment : Switching Mode Power Supply
Model Designation : See Certificate
Certificate No. : TA 50348281 0004
Report No. : 10054494 003

Dear Mr. Frank Lin,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body



Dipl.-Ing. W. Hsu

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan, R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
[http: //www.tuv.com](http://www.tuv.com)

NEW TAIPEI BRANCH:
No. 458-18 & 458-19,
Sec. 2, Fenliao Rd., Linkou Dist.,
New Taipei City 244,
Taiwan, R. O. C.
Tel. (02) 2172-1000
Fax (02) 2172-1322

TAICHUNG BRANCH:
3F., No. 497, Zhongming South Rd.,
West Dist., Taichung City 403,
Taiwan, R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322

KAOHSIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

Certificate



Certificate no.

TA 50348281 04

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Test report no.: ZTW1- 10054494 003

Client Reference: P00158594

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Test Requirement : see above

Change

License Holder : see above

Factory : see above

Remark: 1. The risk management requirements of the standard
were not addressed.

2. This certificate is valid for 5 maximum years.

Appendix: 1

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419036526

TÜV Rheinland Taiwan Ltd
Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)
08/02/2021

Certificate



Certificate no.

TA 50348281 05

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD. HUADU BRANCH
No. 11, Jingu South Road, Huadong Town,
Huadu District, Guangzhou 510890
P.R. China

Test report no.: ZTW1- 10054494 003

Client Reference: P00158594

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

Remark: 1. The risk management requirements of the standard
were not addressed.
2. This certificate is valid for 5 maximum years.

Appendix: 1

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419036526

TÜV Rheinland Taiwan Ltd
Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)
08/02/2021

Certificate



Certificate no.

TA 50348281 06

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Test report no.: ZTW1-LinPr 10054494 004 Client Reference: SC/SPC-2201010-App

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01
Change

Type Designation : RPS-200-x (MW)
x stands for : 12, 15, 24, 27, 30 or 48

Addition

Type Designation : RPS-200-30LIN (MW)

Inputs : see constructional dataform

Outputs : see constructional dataform

Remark: 1. The risk management requirements of the standard
were not addressed.

2. This certificate is valid for 5 maximum years.

Appendix: 2

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419036526

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

18/03/2022

Certificate



Certificate no.

TA 50348281 07

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD.
No. 11,
Jingu South Road, Huadu District,
510890 Guangzhou,
P.R. China

Test report no.: ZTW1-LinPr 10054494 004 **Client Reference:** SC/SPC-2201010-App

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

Remark: 1. The risk management requirements of the standard
were not addressed.
2. This certificate is valid for 5 maximum years.

Appendix: 2

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419036526

TÜV Rheinland Taiwan Ltd

Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

18/03/2022